



United States Department of the Interior



GEOLOGICAL SURVEY
RESTON, VA 22092

REPORT OF CALIBRATION
of Aerial Mapping Camera

September 6, 1990

Camera type: Wild RC10	Camera serial no.: 3562
Lens type: Wild Universal Aviogon /4	Lens serial no.: 13048
Nominal focal length: 153 mm	Maximum aperture: f/4
	Test aperture: f/4

Submitted by: Pacific Aerographics
Santa Ana, California

Reference: Letter dated August 29, 1990, from Mrs. M.A. Andelin.

These measurements were made on Kodak Micro-flat glass plates, 0.25 inch thick, with spectroscopic emulsion type V-F Panchromatic, developed in D-19 at 68° F for 3 minutes with continuous agitation. These photographic plates were exposed on a multicollimator camera calibrator using a white light source rated at approximately 5200K.

I. Calibrated Focal Length: 152.891 mm

This measurement is considered accurate within 0.005 mm

II. Radial Distortion

Field angle	\bar{D}_c	D_c for azimuth angle			
		0° A-C	90° A-D	180° B-D	270° E-C
degrees	um	um	um	um	um
7.5	1	1	1	1	-1
15	0	0	0	0	-1
22.5	-2	-3	0	-2	-1
30	1	2	0	2	0
35	1	2	0	1	1
40	0	0	-1	1	-2

The radial distortion is measured for each of four radii of the focal plane separated by 90° in azimuth. To minimize plotting error due to distortion, a full least-squares solution is used to determine the calibrated focal length. \bar{D}_c is the average distortion for a given field angle. Values of distortion D_c based on the calibrated focal length referred to the calibrated principal point (point of symmetry) are listed for azimuths 0°, 90°, 180° and 270°. The radial distortion is given in micrometers and indicates the radial displacement away from the center of the field. These measurements are considered accurate within 5 um.

III. Resolving Power in cycles/mm

Area-weighted average resolution: 86

Field angle:	0°	7.5°	15°	22.5°	30°	35°	40°
Radial lines	113	113	113	80	95	95	57
Tangential lines	113	95	80	80	95	80	67

The resolving power is obtained by photographing a series of test bars and examining the resultant image with appropriate magnification to find the spatial frequency of the finest pattern in which the bars can be counted with reasonable confidence. The series of patterns has spatial frequencies from 5 to 268 cycles/mm in a geometric series having a ratio of the 4th root of 2. Radial lines are parallel to a radius from the center of the field, and tangential lines are perpendicular to a radius.

IV. Filter Parallelism

The two surfaces of the Wild 525 No. 6097 filter accompanying this camera are within 10 seconds of being parallel. This filter was used for the calibration.

V. Shutter Calibration

<u>Indicated shutter speed</u>	<u>Effective shutter speed</u>	<u>Efficiency</u>
1/200	5.25 ms = 1/190 s	81%
1/400	2.62 ms = 1/380 s	81%
1/600	1.75 ms = 1/570 s	81%
1/800	1.31 ms = 1/760 s	81%
1/1000	1.05 ms = 1/950 s	81%


The effective shutter speeds were determined with the lens at aperture f/4. The method is considered accurate within 3 percent. The technique used is Method I described in American National Standard PH3.48-1972(R1978).

VI. Film Platen

The film platen mounted in Wild RC10 drive unit No. 3562-381 does not depart from a true plane by more than 13 um (0.0005 in).

This camera is equipped with a platen identification marker that will register "381" in the data strip area for each exposure.

VII. Principal Points and Fiducial Coordinates



Positions of all points are referenced to the principal point of autocollimation (PPA) as origin. The diagram indicates the orientation of the reference points when the camera is viewed from the back, or a contact positive with the emulsion up. The data strip is to the left.

	<u>X coordinate</u>	<u>Y coordinate</u>
Indicated principal point, corner fiducials	-0.003 mm	-0.001 mm
Indicated principal point, midside fiducials	-0.002	0.001
Principal point of autocollimation	0.0	0.0
Calibrated principal point (point of symmetry)	-0.004	0.005

Fiducial Marks

1	-106.012 mm	-105.998 mm
2	106.003	105.994
3	-105.990	105.997
4	105.982	-105.998
5	-110.004	0.004
6	109.999	-0.003
7	0.012	110.002
8	-0.017	-110.004

VIII. Distances Between Fiducial Marks

Corner fiducials (diagonals)

1-2: 299.818 mm 3-4: 299.790 mm

Lines joining these markers intersect at an angle of 90° 00' 00"

Midside fiducials

5-6: 220.004 mm 7-8: 220.006 mm

Lines joining these markers intersect at an angle of 89° 59' 40"

Corner fiducials (perimeter)

1-3: 211.995 mm 2-3: 211.993 mm

1-4: 211.994 mm 2-4: 211.992 mm

The method of measuring these distances is considered accurate within 0.005 mm.


IX. Stereomodel Flatness

Drive unit No.: 3562-381

Base/Height ratio: 0.6

Platen ID: 381

Maximum angle of field tested: 40°



Stereomodel
Test point array
(values in micrometers)

The values shown on the diagram are the average departures from flatness (at negative scale) for two computer-simulated stereomodels based on comparator measurements on contact glass (Kodak Micro-flat) diapositives made from Kodak 2405 film exposures. These measurements are considered accurate within 5 μ m.

X. Resolving Power in cycles/mm

Area-weighted average resolution: 40

Film: Type 2405

Field angle:	0°	7.5°	15°	22.5°	30°	35°	40°
Radial lines	57	48	48	40	48	48	40
Tangential lines	57	40	40	34	40	34	28

This aerial mapping camera calibration report supersedes the previously issued USGS Report No. OSL/1289, dated October 21, 1987.

Eberhard G. Schirmacher
Eberhard G. Schirmacher
Chief, Optical Science Laboratory
National Mapping Division